PH-59

41924-000

CT-DI Phantom (Head and Body Phantom)





A set of phantoms for CTDI-100, conforming to requirements described in 21 CFR 1020.33, IEC 61223-3-5: 2004, and IEC 61223-2-6: 2006 as consistency test







A set with different type of tissue substitute can be custom-ordered

FEATURES

I Represent adult head and body as well as pediatric body Can be used for initial and follow-up QA tests

Evaluation Parameters

Computed Tomography Dose Index (CTDI) I Dose profile

DESCRIPTIONS

SET INCLUDES

- 1 head phantom
- body phantom
- filling rods manual

SPECIFICATIONS

Phantom size:

Body phantom: 32 dia. x 15 cm / 12.6 dia. x 5.9 in Head phantom: 16 dia. x 15 cm/ 6.3 dia. x 5.9 in

Phantom weight: 15kg/ 33lb 4kg/ 8.8lb

MATERIALS Acrylic resin

PH-54

41919-000


CT QA Phantom JCT Ⅱ






Renovated CT QA phantom for initial test at the time of reception, as well as follow-up periodical quality check of CT scanners









FEATURES

The phantom can be used for initial and follow-up QA tests listed below, described in JIS Z 4752-3-5: 2008 (IEC 61223-3-5: 2004) and Z 4752-2-6: 2012 (IEC 61223-2-6: 2006)

| Conforming to JIS Z 4923:2015











1 vaseline

manual

screws (spare)

1 angle adjustment holder

| Axial scan:

Slice thickness / spatial resolution / low contrast resolution / noise / mean HU number / uniformity

EVALUATION PARAMETERS

I Helical scan: Slice thickness

System requirements of software : OS Windows 7 (64 bit), Windows 8.1 Pro (64 bit), memory 4GB, HDD 250 GB»

DESCRIPTIONS

SET INCLUDES

- 1 cylindrical container (fixing screw)
- slice thickness unit (axial)
- spatial resolution unit
- repeated pattern unit
- 1 low contrast resolution unit
- slice thickness unit (helical)
- 1 fixture for the slice thickness unit fixture for the cylindrical container

Phantom weight:

MATERIALS

Acrylic resin, polyurethane, stainless

SPECIFICATIONS Phantom size: 20 dia. x 20 cm / 7.9 dia. x 7.9 in

3 kg / 6.6 lb



